



PO Box 5183
Oak Ridge, TN 37831
(865) 999-7873
info@Rn86Solutions.com

Radon Test Report

July 08, 2025
Batch #: 070825-1
Invoice #:

Customer:

Herman & Lily Munster
1313 Mockingbird Lane
Knoxville TN 37913

Test Site:

Herman & Lily Munster
1313 Mockingbird Lane
Knoxville TN 37913

Rad Elec Recon Continuous Radon Monitors were used for radon screening measurements that were conducted at the above referenced test site by:

The Results are as follows:

Serial	Type	Location	Test Start Date	Test End Date	Results (pCi/L)
415	CRM	Crypt	06-Feb-2025 5:00 PM	08-Feb-2025 5:05 PM	43.2


CRM Calibration Due: 27-Jan-2026

Average Radon Concentration in: Crypt **43.2 pCi/L**

Deployed By: Lee Ruble NRPP 111871-RMP
Retrieved By: Lee Ruble NRPP 111871-RMP
Analyzed By: Lee Ruble NRPP 111871-RMP

Radon Health Risk Information

Radon is the second leading cause of lung cancer after smoking. The U.S. Environmental Protection Agency (EPA) and the Surgeon General strongly recommend that further action be taken when a home's radon test results are 4.0 pCi/L or greater. The national average indoor radon level is about 1.3 pCi/L. The higher the home's radon level, the greater the health risk to you and your family. Reducing your radon levels can be done easily, effectively and fairly inexpensively. Even homes with very high radon levels can be reduced below 4.0 pCi/L. Please refer to the EPA website at www.epa.gov/radon for further information to assist you in evaluating your test results or deciding if further action is needed.

Signature: 

Date: 12-Jul-2025



PO Box 5183
 Oak Ridge, TN 37831
 (865) 999-7873
 info@Rn86Solutions.com

Radon Test Report

July 08, 2025
 Batch #: 070825-1
 Invoice #:


Customer:

Herman & Lily Munster
 1313 Mockingbird Lane
 Knoxville TN 37913

Test Site:

Herman & Lily Munster
 1313 Mockingbird Lane
 Knoxville TN 37913

Humidity & Pressure



Hourly Detailed Report for: Rad Elec Recon 415

Record #	Date/Time	pCi/L	Temperature (°F)	Pressure (inHg)	Humidity (%)	Tilts
1	06-Feb-2025 18:00	15.7	58	28.9	63	0
2	06-Feb-2025 19:00	18.9	58	28.9	63	0
3	06-Feb-2025 20:00	21.7	58	28.9	63	0
4	06-Feb-2025 21:00	22.8	58	28.9	63	0
5	06-Feb-2025 22:00	23.5	58	29.0	63	0
6	06-Feb-2025 23:00	22.5	58	29.0	63	0
7	07-Feb-2025 0:00	21.0	58	29.0	63	0
8	07-Feb-2025 1:00	26.9	58	29.1	63	0
9	07-Feb-2025 2:00	24.6	58	29.1	63	0
10	07-Feb-2025 3:00	29.8	58	29.1	64	0
11	07-Feb-2025 4:00	27.2	58	29.1	64	0
12	07-Feb-2025 5:00	29.0	58	29.1	64	0
13	07-Feb-2025 6:00	28.4	58	29.2	64	0
14	07-Feb-2025 7:00	34.5	58	29.2	64	0
15	07-Feb-2025 8:00	32.7	58	29.2	63	0
16	07-Feb-2025 9:00	30.5	58	29.2	63	0
17	07-Feb-2025 10:00	33.7	58	29.2	63	0
18	07-Feb-2025 11:00	36.2	58	29.2	62	0
19	07-Feb-2025 12:00	33.8	57	29.2	62	0
20	07-Feb-2025 13:00	34.4	57	29.2	61	0
21	07-Feb-2025 14:00	35.7	57	29.1	61	0
22	07-Feb-2025 15:00	35.6	57	29.1	61	0
23	07-Feb-2025 16:00	33.0	57	29.1	61	0
24	07-Feb-2025 17:00	31.4	57	29.1	61	0
25	07-Feb-2025 18:00	39.3	57	29.1	61	0



PO Box 5183
 Oak Ridge, TN 37831
 (865) 999-7873
 info@Rn86Solutions.com

Radon Test Report

July 08, 2025

Batch #: 070825-1

Invoice #:

Customer:

Herman & Lily Munster
 1313 Mockingbird Lane
 Knoxville TN 37913

Test Site:

Herman & Lily Munster
 1313 Mockingbird Lane
 Knoxville TN 37913

Record #	Date/Time	pCi/L	Temperature (°F)	Pressure (inHg)	Humidity (%)	Tilts
26	07-Feb-2025 19:00	38.2	57	29.1	61	0
27	07-Feb-2025 20:00	37.8	57	29.1	61	0
28	07-Feb-2025 21:00	40.6	57	29.1	61	0
29	07-Feb-2025 22:00	42.0	57	29.1	61	0
30	07-Feb-2025 23:00	44.7	57	29.1	61	0
31	08-Feb-2025 0:00	51.4	57	29.1	61	0
32	08-Feb-2025 1:00	50.6	57	29.1	61	0
33	08-Feb-2025 2:00	54.1	57	29.1	61	0
34	08-Feb-2025 3:00	51.4	57	29.1	61	0
35	08-Feb-2025 4:00	55.2	57	29.1	61	0
36	08-Feb-2025 5:00	55.7	57	29.1	61	0
37	08-Feb-2025 6:00	58.4	57	29.1	61	0
38	08-Feb-2025 7:00	63.1	57	29.0	61	0
39	08-Feb-2025 8:00	63.0	57	29.1	61	0
40	08-Feb-2025 9:00	64.6	57	29.1	61	0
41	08-Feb-2025 10:00	69.0	57	29.1	61	0
42	08-Feb-2025 11:00	67.2	57	29.1	61	0
43	08-Feb-2025 12:00	74.7	57	29.0	61	0
44	08-Feb-2025 13:00	70.2	57	29.0	61	0
45	08-Feb-2025 14:00	76.9	57	29.0	61	0
46	08-Feb-2025 15:00	72.5	57	28.9	61	0
47	08-Feb-2025 16:00	77.7	57	28.9	62	0
48	08-Feb-2025 17:00	73.6	57	28.9	62	0